

Report for:

Pat Hartshorn - Advance Look Building Inspections & Environmental
RespirCare Analytical Network
US Division
Morristown, AZ 85342


Regarding:

Project #1815 Momeganna St, Airport Inn
EML ID: 3285799

Spore trap analysis (EM-MY-S-1038)
AIHA-LAP, LLC accredited service, Lab ID #102297
1501 West Knudsen Drive
Phoenix, AZ 85027

Dates of Analysis:
Spore trap analysis 06-22-2023

Approved by:



Business Unit Manager
Joshua Cox

Quality Control (QC) Information

Secondary data review and release completed¹ : 06-22-2023 5:12PM MDT

Last QC sample completed² : 06-22-2023 5:11PM MDT

of projects since last QC sample completed³ : 0 (this project)

Analyst daily reference slide completed⁴ : Yes

Daily reference analysis#: 15836765 **Completed:** 06-22-2023 8:16AM MDT

Eurofins EMLab P&K's LabServe® reporting system includes automated fail-safes to ensure that all AIHA-LAP, LLC quality requirements are met and notifications are added to reports when any quality steps remain pending.

1. AIHA-LAP, LLC policies require that all data is reviewed prior to reporting by a second qualified individual unless noted on the report. A "Delayed" entry indicates that the secondary review is still pending, and data is considered preliminary. See individual sample comments for more information.
2. This timestamp refers to the last Quality Control sample completed at the given facility for the given service type.
3. This entry refers to the number of projects with analyses completed for this service type by this analyst since the last quality control sample evaluation for the analyst. A result of zero indicates that a sample from this project was analyzed for Quality Control purposes. AIHA-LAP, LLC policies require the repeat analysis of 10% of analyst samples for Quality Control, which Eurofins EMLab P&K applies to batches of 20 samples.
4. AIHA-LAP, LLC policies require that all analysts must successfully complete a daily reference slide prior to analyzing any samples for this service type on a given day.